



SAFESEPT^{MAX} Instrument Cleaner Enzymatic Concentrate

- strong cleaning power
- ideal for ultrasonic cleaning
- excellent material compatibility



PRODUCT DESCRIPTION:

SAFESEPT MAX® Instrument Cleaner Enzymatic Concentrate offers an excellent cleaning performance due to the combination of highly surface active nonionic surfactants and a stabilised triple enzyme complex.

SAFESEPT MAX® Instrument Cleaner Enzymatic Concentrate offers an excellent material compatibility and is therefore the ideal cleaning concentrate for manual reprocessing of surgical instruments and flexible endoscopes. It is economic in use and recommended for use in ultrasonic baths.

The self-acting formula counteracts, penetrates and effectively removes all kinds of organic contaminations such as proteins, coagulated blood, mucus, vomit, fat and fecal matter even in hard-to-reach areas.

AREAS OF APPLICATION:

Highly-efficient, enzymatic cleaning concentrate for manual instrument reprocessing. Suitable for ultrasonic baths.

CONCENTRATION:

Instrument cleaning: 1% (10 ml/L).

Cleaning in ultrasonic bath: 0.5% (5 ml/L).

Temperature: Mix solution with warm water (30 - 40°C).

Exposure time: At least 5 minutes.

APPLICATION:

Immerse instruments and endoscopes into the cleaning solution and / or ensure a complete wetting of all instrument surfaces and cavities. Completely remove air from lumens. Brush instruments if necessary. Clean endoscope channels mechanically. Remove cleaned instruments and rinse thoroughly with clear water prior to further reprocessing steps.

COMPOSITION:

5 - 15% Tensioactivos no iónicos, 2-metil-2H-iso-tiazol-3-ona, 1,2-bencisotiazol-3(2H)-ona, Enzimas Proteasa, amilasa, lipasa.



SAFESEPT^{MAX} Instrument Cleaner Enzymatic Concentrate

- strong cleaning power
- ideal for ultrasonic cleaning
- excellent material compatibility



SPECIAL INFORMATION:

Can cause serious eye damage. Wear protective gloves/protective clothing/eye protection/face protection. IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER/doctor.



In case of visible soiling the SAFESEPT MAX[®] Instrument Cleaner Enzymatic solution must be changed immediately; in normal circumstances, change it at least every working day.

PRODUCT STATUS/LISTINGS:

- Conform to the EU guideline 93/42/EC for medical devices.

AVAILABLE DELIVERY FORMS:

REF **988-2764** 1L Dosing Bottle